



Product family: BLV
Document Nr: CE 900080-001

We,

August 9, 2012

Freedom Scientific, Inc.
11800 31st Court North, St. Petersburg, FL 33716 USA
With authorized EU Representative Freedom Scientific GmbH
High-Tech Center 2, Bahnstrasse 1, 8274 Tagerwielen Switzerland

as manufacturer declare under our sole responsibility that the following described equipment in our delivered version

Product Names:

Focus Braille Displays

Model Names:

Focus 40B

Focus 14B

complies with and is in accordance with the following Directives and carried the CE Marking accordingly:

- EMC Directive 2004/108/EEC
- Low Voltage Directive 2006/95/EEC
- Medical Directive 93/42/EEC

CE:

IEC Standards and Directives:

- | | |
|--------------------------------------|---------------------------------------|
| Medical Directive 93/42/EU Class 1 | |
| EN 55022:2006 | EN 61000-2:2006 |
| EN 55024:1998 / A1:2001 / A3:2003 | EN 61000-3-3:1995 / A1:2001 / A2:2005 |
| EN 61000-3-2:2006 | EN ISO 14971 |
| EN 61000-4-2:1995 | EN 12182 |
| EN 61000-4-3:2002 / A1:2002 | UL 60950-1:2007 |
| EN 6100-4-4:1995 / A1:2001 / A2:2001 | EN 60950-1 / A12:2011 |
| EN 61000-4-5:1995 | RoHs 2002 / 95 / EC |
| EN 610000-4-6:1996 | |
| EN 61000-4-8:1994 | |
| EN 61000-4-11:1994 | |

Authorized Signature / Date

CFO

8/10/12

Manufacturer or Authorized Representative in EU

Title

Date